



## APPLICATION

NYU LangoneHealth

## PURPOSE

To establish guidelines for the safe use of cellular and wireless technology

## POLICY

### 1.0 Application

NYU Langone Health (NYULH) refers to the NYU Langone Health System, NYU Langone Hospitals, NYU Grossman School of Medicine, NYU Long Island School of Medicine, the Family Health Centers at NYU Langone, and all entities controlled by any of them.

This policy applies to:

All indoor and outdoor areas of all NYULH owned and leased facilities

All employees, contractors, and consultants of NYULH

All patients, visitors, and guests of NYULH

### 2.0 Policy

2.1 Cellular phone usage is permitted in all areas of the NYULH with the following caution

- x Users shall connect to WiFi provided by NYULH
- x Users shall maintain a distance of at least 1 meter from medical equipment as recommended by ECR, when practical, while using a mobile communication device
- x Users shall avoid placing communication devices on top of, or in direct contact with any medical device
- x Users shall be cognizant of patient privacy and to avoid distraction or unnecessary noise while using mobile devices
- x Staff shall monitor energized medical devices, and report all incidents involving EMI interference of medical devices to Clinical Engineering promptly for evaluation. See Appendix A for Supplemental Educational Materials.

2.2 Air waves are unprotected and conversations may be intercepted by other telephones. Staff will refrain from disclosing protected health information during cellular phone conversations to protect patient confidentiality.



2.3 Two-way Radios may be used in the receive mode only in all patient areas.

2.4 Emergency Use of Two-Way Radios

- x Whenever maintenance staff, security officers, EMS personnel, Life Flight employees, or other personnel with two-way radios are in a patient care unit and an emergency occurs which requires the use of their radio, they should make their best effort to adhere to the following guidelines, listed in preferred order:
  - x Leave the patient care unit to use the two-way radio. (This minimizes interference created by RF devices as the distance between the RF device and the affected equipment is increased.)
  - x Keep the two-way radio at least ten (10) feet away from any energized medical device.
  - x If output levels are adjustable, use the lowest setting possible that facilitates acceptable communications.
  - x If any equipment in the vicinity of the radio user should malfunction while the radio is in use, terminate use of the radio immediately. Any further use should be from the visitor waiting area or a non-patient care unit.

2.5 Tablets, Laptops and Notebooks must undergo a doc testing against the major medical equipment to ensure that they do not cause any interference.

2.6 Special Cases

- x Implanted Devices Patients and staff with implanted devices such as pacemakers and defibrillators, must exercise caution when carrying and using wireless devices. Recommended procedures, as delineated by the medical device manufacturer must be followed.
- x Emergency Vehicles Operators of emergency vehicles, which transport patients with sensitive medical devices, and the processors, must be cognizant of the fact that two-way radios, beeper and cell phones can cause electromagnetic interference at close range. Medical devices used in these applications must be carefully selected.

2.8 Exceptions to this policy must be approved by the Clinical Safety Committee.

2.9 Any questions pertaining to this policy should be directed to Clinical Engineering

3.0 Incident Reporting

3.1 Report any suspected incidents of medical device electromagnetic interference to Clinical Engineering. Clinical Engineering shall investigate each incident,



develop recommendations, and report findings to the Patient Safety Officer, Risk Management, the Clinical Safety Committee and/or the Environment of Care (EOC) Committee or Subcommittee as appropriate.

- 3.2 Document patient injuries in Patient Safety Intelligence (PSI). See Administrative Policy for Patient Safety/Adverse Event Reporting Policy and Reporting & Follow-up of Adverse Events/Occurrences (Incident Reporting), available on Ellucid.
- 3.3 Contact Clinical Engineering for any questions about the applicability of using and enforcing the use of wireless devices.

#### 4.0 Responsibilities

- 4.1 Supply Chain is responsible for
  - x I



- x The undisciplined use of cellular devices in the OR—whether for telephone, mail, or data communication, and whether by the surgeon or by other members of the surgical team—may pose a distraction and may compromise patient care.
- x Surgeons should be considerate of the duties of personnel in the OR suite and refrain from engaging them unnecessarily in activities, including assistance in cellular communication, that might divert attention from the patient or the conduct of the procedure.
- x Cellular phones must not interfere with patient monitoring devices or with other technologies required for patient care.
- x Whenever possible, members of the OR team, including the operating surgeon, should only engage in urgent or emergent outside communication during surgery. Personal and routine calls should be minimized. Calls should be kept as brief as possible.
- x Whenever possible, incoming calls should be forwarded to the OR desk or to the hardwired telephone in the OR to minimize the potential distraction of cellular phones.
- x The ring tone should be silenced. An inaudible signal may be employed.
- x Whenever possible, a distinct signal for urgent or emergent calls should be enabled. This signal may be implemented via a “page” option in most cellular telephones. Callers should be advised to use this function only for urgent and emergent calls if the phone is unanswered.
- x The use of cellular devices or their accessories (such as earphones or keyboards) must not compromise the integrity of the sterile field.
- x Special care should be taken to avoid sensitive communication within the hearing of awake or sedated patients.
- x Communication using hardwired phones in the operating room is subject to the same discipline as communication using cellular technology.
- x The use of cellular devices to take and transmit photographs should be governed by hospital policy on photography of patients and by government regulations pertaining to patient privacy and confidentiality.